

OCT Assessment:

1. What is the pullback length of the "Survey mode"?

- A. 71mm
- B. 73mm
- C. 75mm
- D. 77mm

2. What is the approved injector pump settings relating to injecting contrast medium / pressure limit/ rise time?

- A. 100%/ 300psi/ 0.5sec
- B. 50%/ 400psi/ 0sec
- C. 100%/ 600psi/ 0sec
- D. 100%/ 300psi/ 0sec

3. How many radiopaque markers on OCT catheter can be seen under angiogram?

- A. 1
- B. 2
- C. 3
- D. 4

4. What should we do before putting DOC into DOC cover?

- A. clean the DOC
- B. open dust cap
- C. connect purge syringe
- D. connect OCT catheter

5. If EEL can be visualize in Distal reference, we should round up to the nearest stent size

- A. True
- B. False